

# CLASS VIII ALS CHECKLIST



Agency Name \_\_\_\_\_  
VIN (Last 4) \_\_\_\_\_

License # \_\_\_\_\_  
Unit # \_\_\_\_\_

VEHICLE PERMIT # \_\_\_\_\_  
Date \_\_\_\_\_

AGENCY OPERATIONS	
	Maintain two (2) complete sets of equipment
	Each basic life support non-transport vehicle shall wrap, properly store, and handle all single-service implements to be inserted into the patient's nose or mouth.
	Shall wrap, properly store, and handle all single-service implements to be inserted into the patient's nose or mouth.
	Shall properly store and keep multiuse items clean and sterile if indicated.
	Capability to communicate on all VHF Ky State Mutual Aid Frequencies, per the Ky Field Operations Guide
SUCTION, VENTILATION EQUIPMENT	
	Capable of minimum vacuum of 300mm Hg
	Wide-bore tubing
	Rigid catheters
	Soft pharyngeal suction tips in child size
	Soft pharyngeal suction tips in adult size
	Bag Valve Mask in following sizes:
	Adult
	Child
	Infant
	Neonatal mask only
	Two (2) oropharyngeal airways in each of the following sizes:
	Adult
	Child
	Infant
	One (1) pocket mask with a one (1) way valve
	Blind-Insertion Airway Devices (BIAD) in adult and pediatric sizes
OXYGEN EQUIPMENT	
	Portable oxygen equipment of at least 300 liters capacity and D size cylinder with a regulator capable of delivering 25LPM

	Two (2) non-rebreathing oxygen masks in both adult and pediatric sizes
	Two (2) nasal cannula in both adult and pediatric sizes
	Two (2) nasopharyngeal airways with water-soluble lubricant in each of the following sizes:
	Adult
	Child
	Infant
BANDAGING SUPPLIES	
	Two (2) airtight dressings for open chest wounds
	Assorted bandaging supplies for the care of soft tissue injuries
	Sterile water for irrigation
	An AED with a minimum of two (2) complete sets of pads for all non-ALS providers and vehicles
SPLINTING & IMMOBILIZATION	
	Two (2) upper extremity-splinting devices
	Two (2) lower extremity-splinting devices
	Two (2) cervical collars in each of the following sizes or adjustable equivalents:
	Pediatric
	Small
	Medium
	Large
	No-Neck
PERSONAL PROTECTIVE EQUIPMENT FOR EACH STAFF MEMBER	
	Gown
	Face mask and shield
	Gloves
	Biohazard bag
	Puncture resistant container for the disposal of sharp objects

	Antimicrobial hand cleaner
<b>OB SUPPLIES</b>	
	Sterile scalpels or scissors
	Sterile gloves
	Bulb suction
	Two (2) umbilical clamps
<b>BLOOD PRESSURE EQUIPMENT</b>	
	One (1) blood pressure sphygmomanometer in each of the following cuff sizes:
	Large adult
	Adult
	Pediatric
	One (1) stethoscope in each of the following sizes:
	Adult
	Pediatric
	A glucometer or blood glucose-measuring device with reagent strips and lancets for obtaining blood glucose sample.
<b>MISCELLANEOUS EQUIPMENT</b>	
	Eye protection goggles or safety glasses
	Heavy work gloves
	Hard hats that meet ANSI standards, as stated in 29 C.F.R. 1910.135
	Reflective safety wear for each crew member that meet current ANSI standard ANSI 107-2010 or ANSI 207-2011
	Three (3) reflective triangles or strobes, or equivalent warning devices.
	Pediatric drug dosage tape or equivalent that provides easy reference for pediatric and infant treatment and drug dosages.
	Controlled drugs shall be stored in a locked storage box in a locked compartment that is immediately accessible to personnel.

<b>ALS EQUIPMENT</b>	
<b>ENDOTRACHEAL INTUBATION EQUIPMENT</b>	
	Laryngoscope handle
	Various laryngoscope blades in adult, pediatric, and infant sizes
	Extra batteries and bulbs for handles or blades
	A minimum of seven (7) different sizes of endotracheal tubes for oral and nasal placement in adult, pediatric, and infant sizes
	Equipment necessary to perform emergency cricothyrotomy
	An end tidal carbon dioxide detection device
	Stylettes in adult and pediatric sizes
	Magill forceps in adult and pediatric sizes
	One-half (1/2) inch wide twill tape or equivalent for securing endotracheal tubes
	Water-soluble lubricant for lubrication of endotracheal and nasotracheal tubes
<b>MEDICATION/IV ADMINISTRATION</b>	
	Sterile, disposable needles, in types and sizes sufficient for personnel to administer medications and perform procedures allowed by the agency's patient treatment protocols

Disposable syringes in types and sizes sufficient for personnel to administer medications and perform procedures allowed by the agency's patient treatment protocols
Restriction band appropriate for use with venipuncture procedure
Disposable, individually packaged antiseptic wipes
Intravenous fluids as required by the agency's protocol, with macrodrip and microdrip fluid sets, and accessory items including over the needle catheter devices in sizes fourteen (14) to twenty-four (24) gauge
Intraosseous needles or intraosseous devices appropriate for children and adults
Sterile, disposable needles, in types and sizes sufficient for personnel to administer medications and perform procedures allowed by the agency's patient treatment protocols
Disposable syringes in types and sizes sufficient for personnel to administer medications and perform procedures allowed by the agency's patient treatment protocols
Restriction band appropriate for use with venipuncture procedure
Disposable, individually packaged antiseptic wipes

Intravenous fluids as required by the agency's protocol, with macrodrip and microdrip fluid sets, and accessory items including over the needle catheter devices in sizes fourteen (14) to twenty-four (24) gauge
<b>MONITOR/DEFIBRILLATOR</b>
Capable of displaying a visual display of cardiac electrical activity
Capable of providing a hard copy of cardiac electrical activity measure
Capable of delivering direct current energy over a variable range, which is suitable for pediatric and adult usage
Capable of providing external cardiac pacing
Has adult and pediatric external pads, capable of utilization for immediate monitoring of heart activity and delivery of counter shock in both the adult and pediatric patient
Is capable of being operated from internal rechargeable batteries
Has synchronized counter shock capability for cardioversion
Has a patient monitoring cable with electrode pads or equivalent for use with the patient monitoring cable

